

# RETURN AND REPAIR FORM



## Customer Information

B:

S:

Company Name: Franklin Twp. P.D., Somerset, NJ

Date Received: 06/13/13

Date given to service: 06/13/13

Carrier:  FedEx  UPS  USPS Shipping Method:  GRD  3DAY  2DAY  
 NDA-PRI  NDA-STD

Product:  6510  6810  
 7110  8610  
 7510  DT5000  
 7410 Upper-half  9510  
 7410 Whole

Serial Number: AR WA-0166

Printer Serial#: AR

Sim Serial#: DD

Probe Serial#: DD

Warranty Expires: \_\_\_\_\_

Description:  A  B  Plus  Demo  Screener  Trade In

## Accessories

110V A/C Adapter  Regulator  Printer Ribbon  Printer Paper  
 Mouthpieces  9510 Stylus  9510 Top Cover  Carrying Case  
 Dry Gas  Other (please specify) \_\_\_\_\_

## Repair Information

Test#: \_\_\_\_\_

| Part Number | Description   | Qty | Total Cost |
|-------------|---------------|-----|------------|
| MP Cal 71   | Calibration   | 1   | NC/W       |
| 6808455     | Fuel Cell     | 1   | NC/W       |
| 6808486     | Spacer Plates | 2   | NC/W       |
| MP Labor    | Labor         | .5  | NC/W       |

Repair Notes: Replaced defective fuel cell and spacer plates.

Fuel Cell = ARED-0139

CAL W/QC AND OPS CHECK

Service Technician BC

Date: 06/24/13