

Dräger safety

Return & Repair Form

Customer Information:

CUSTOMER Waterford P.D. N.J. Cust # _____
 Date of Return: 10/24/06 Carrier: FedEx Method: Priority

Product Information:

Product : (7110) 7410 6510 6810 Serial # ARWA-0170
 Description: A B PLUS SCREENER _ Printer Serial # _____
 Whole Inst. Top Half Simulator Serial # _____
 OTHER : _____ Probe Serial # _____
 Demo Unit Returned to stock on: _____
 Accessories: (check all that apply):
 110VAC Adapter Regulator Mag Card Rdr # _____
 Printer Paper Printer Ribbon Casio # _____
 Mouthpieces Carrying Case Dry Gas
 Other (specify): _____
 Warranty Expiration Date: _____

Repair Information:

Test # _____

Reason for Return: Cal

Part #	Description	Qty	Total
<u>6809512</u>	<u>Hose barb w/ check valve</u>	<u>1</u>	
<u>6808455</u>	<u>Fuel Cell</u>	<u>1</u>	
<u>MPLABOR</u>	<u>LABOR</u>		
<u>MPALTI</u>	<u>7110 CAL</u>	<u>1</u>	

Repair Notes: Resoldered EC motor connections due to cold solder joint. Replaced sticky check valve. Replaced possibly leaky F/C. Full cal ops check + Q.C.

Technician: CRD

Date: 10/25/06