

# Return & Repair Form:

## Customer Information:

Company Name: Clark P.D. B: \_\_\_\_\_ S: \_\_\_\_\_  
 Date Received: 2-16-10 Date Given to Service: \_\_\_\_\_  
 Carrier: FedEx - UPS - DHL - USPS Method: GRD - NDA - 2DY - 3DY - Other

## Product Information:

Product: 7410-~~7110-6510-6810-7510~~  
 9510- Drug Tester 5000 Serial #: AR WA-0173  
 Description: A - B - Plus - Screener - Demo Printer Ser #: AR -  
 Whole Instrument Top 1/2 Sim Ser #: \_\_\_\_\_  
 Other \_\_\_\_\_ Probe Ser # DD P -  
**ACCESSORIES:**  
 110 V A/C Adapter  Regulator  Mag Card Rdr# \_\_\_\_\_  
 Printer Paper  Printer Ribbon  Casio # \_\_\_\_\_  
 Mouthpieces  Carrying Case  Dry Gas  
 Other (Please Specify) \_\_\_\_\_ Warranty Exp. Date \_\_\_\_\_

## Repair Information:

Test #

Reason for Return: "NO ADMITTANCE" FOR BLACK KEY?

Part Number	Description	Qty	Total Cost
<u>MPLABTR</u>		<u>.5</u>	<u>W/CWTR</u>

## Repair Notes:

PROBLEM NOT DUPLICATED  
CHECKED CIRC. VIA SERVICE CH 3 = ALL OK  
CHECKED AND GAINED ACCESS TO BLACK KEY FUNCTIONS 50+  
TIMES = NO PROBLEMS

Service Technician: [Signature]

Date: 02-19-10