

RETURN AND REPAIR FORM



Customer Information

B:

S:

Company Name: Hillsborough TWP P.D., Hillsborough, NJ

Date Received: 02/18/13 Date given to service: 02/18/13

Carrier: FedEx UPS USPS Shipping Method: GRD 3DAY 2DAY
 NDA-PRI NDA-STD

Product: 6510 6810
 7110 8610
 7510 DT5000
 7410 Upper-half 9510
 7410 Whole

Serial Number: AR WA-0188
 Printer Serial#: AR
 Sim Serial#: DD
 Probe Serial#: DD

Warranty Expires: January 2014

Description: A B Plus Demo Screener Trade In

Accessories

110V A/C Adapter Regulator Printer Ribbon Printer Paper
 Mouthpieces 9510 Stylus 9510 Top Cover Carrying Case
 Dry Gas Other (please specify) _____

Repair Information:

Test#:

| Part Number | Description | Qty | Total Cost |
|-------------|--------------|-----|------------|
| MP Cal 71 | Calibration | 1 | NCW |
| 6808455 | Fuel Cell | 1 | NCW |
| 6808486 | Spacer Plate | 1 | NCW |
| MP Labor | Labor | .5 | NCW |

Repair Notes: Replaced defective fuel cell and spacer plate.

Fuel Cell ARDM-2238

CAL W/QC AND OPS CHECK

Service Technician BC Date: 02/22/13