

Return & Repair Form:

Customer Information:

Company Name: Bridgeton City B: _____ S: _____

Date Received: 3-23-09 Date Given to Service: _____

Carrier: RedEx - UPS - DHL - USPS Method: GRD - NDA - 2DY - 3DY - Other

Product Information:

Product: 7410 - 7110 - 6510 - 6810

Serial #: AR WA-0190

Description: A - B - Plus - Screener - Demo

Printer Ser #: AR _____

Whole Instrument

Top 1/2

Sim Ser #: _____

Other _____

Probe Ser # DD P _____

ACCESSORIES:

110 V A/C Adapter

Regulator

Mag Card Rdr# _____

Printer Paper

Printer Ribbon

Casio # _____

Mouthpieces

Carrying Case

Dry Gas

Other (Please Specify) _____

Warranty Exp. Date _____

Repair Information:

Test #

Reason for Return:

Part Number	Description	Qty	Total Cost
6808455	FC	1	N/C WARR
6808486	PLATE	1	N/C WARR
MPCAL 71	CR	1	N/C WARR
MPCAL BCR		.5	N/C WARR

Repair Notes:

REPLACED FC + PLATE

CAL W/ QC & OPS CHECKS

Service Technician: AS/AE

Date: 03/25/2009