

Return & Repair Form:

Customer Information:

Company Name: Fort Dix 000 Joint Base P.O. B: _____ S: _____
 Date Received: 3-4-10 Date Given to Service: _____
 Carrier: FedEx - UPS - DHL - USPS Method: GRD - NDA - 2DY - 3DY - Other

Product Information:

Product: 7410-7110-6510-6810-7510
 9510- Drug Tester 5000
 Description: A - B - Plus - Screener - Demo
 Serial #: AR WA-0194
 Printer Ser #: AR -
 Whole Instrument Top 1/2 Sim Ser #: _____
 Other _____ Probe Ser # DD P -
ACCESSORIES:
 110 V A/C Adapter Regulator Mag Card Rdr# _____
 Printer Paper Printer Ribbon Casio # _____
 Mouthpieces Carrying Case Dry Gas
 Other (Please Specify) _____ Warranty Exp. Date _____

Repair Information:

Test #

Reason for Return:

FC

Part Number	Description	Qty	Total Cost
<u>6808455</u>	<u>FC</u>	<u>1</u>	<u>N/C WARR</u>
<u>6808486</u>	<u>Plate</u>	<u>1</u>	<u>N/C WARR</u>
<u>MP CAL 71</u>	<u>Cal</u>	<u>1</u>	<u>N/C WARR</u>
<u>MP Cabo</u>	<u>Cabo</u>	<u>.5</u>	<u>N/C WARR</u>

Repair Notes:

new FC # ARAN-0898

cal w/Pl, ops check

Service Technician: JS

Date: 03/04/2010