

Return & Repair Form

Customer Information:

Company Name: NJSP B: _____ S: _____
 Date Received: 3/26/09 Date Given to Service: 3/27/09
 Carrier: FedEx - UPS - DHL - USPS Method: GRD - NDA - 2DY - 3DY - Other

Product Information:

Product: 7410 - 7110 - 6510 - 6810 Serial #: AR WC - 0003
 Description: A - B - Plus - Screener - Demo Printer Ser #: AR -
 Whole Instrument Top 1/2 Sim Ser #: _____
 Other _____ Probe Ser # DD P -
ACCESSORIES:
 110 V A/C Adapter Regulator Mag Card Rdr# _____
 Printer Paper Printer Ribbon Casio # _____
 Mouthpieces Carrying Case Dry Gas
 Other (Please Specify) _____ Warranty Exp. Date _____

Repair Information:

Test #

Reason for Return: ERROR 112 + FIX DISPLAY FC

Part Number	Description	Qty	Total Cost
<u>1211</u>	<u>T-FUSE</u>	<u>1</u>	<u>N/C WORK</u>
<u>Q808455</u>	<u>FC</u>	<u>1</u>	<u>N/C WORK</u>
<u>8315075</u>	<u>PCB</u>	<u>1</u>	<u>N/C WORK</u>
<u>MPCAL71</u>	<u>CAL</u>	<u>1</u>	<u>N/C WORK</u>
<u>MP Cal 71</u>	<u>Labor</u>	<u>15</u>	<u>N/C WORK</u>

Repair Notes:

RESET ERROR 112 - MULTIPLE ERRORS DISPLAYED DURING
QC - REPLACE MAIN PCB - T FUSE BLOWN -
REPLACE FC (ABS SEALS)
CAL w/ QC & OPS CHECKS
GA

Service Technician: [Signature]
 Created on 4/10/2008 7:22 AM

Date: 04/22/2009