

# Drägersafety

Internal Use Only

## Return & Repair Form

### Customer Information:

CUSTOMER Tewksbury Twp. Police Cust # \_\_\_\_\_  
Date of Return: 10/19/06 Carrier: FED EX Method: Priority

### Product Information:

Product: 7110 7410 6510 6810 Serial # ARWC-0006  
Description: A B PLUS SCREENER \_ Printer Serial # \_\_\_\_\_  
Whole Inst. Top Half Simulator Serial # \_\_\_\_\_  
OTHER: \_\_\_\_\_ Probe Serial # \_\_\_\_\_  
Demo Unit  Returned to stock on: \_\_\_\_\_  
Accessories: (check all that apply):  
 110VAC Adapter  Regulator  Mag Card Rdr # \_\_\_\_\_  
 Printer Paper  Printer Ribbon  Casio # \_\_\_\_\_  
 Mouthpieces  Carrying Case  Dry Gas  
 Other (specify): \_\_\_\_\_  
Warranty Expiration Date: \_\_\_\_\_

### Repair Information:

Test # \_\_\_\_\_

Reason for Return: EC drops out on 2nd out of 3 tests

Part #	Description	Qty	Total
<u>MPCAL71</u>	<u>7110 CAL</u>	<u>1</u>	

Repair Notes: Could not duplicate stated problem. Tried to induce problem by creating seal leak in simulator w/o success. Performed Full cal ops check + Q.C.

Technician: CRD

Date: 10/27/06