

Return & Repair Form:

Customer Information:

Company Name: NTSP B: _____ S: _____
 Date Received: 10-23-09 Date Given to Service: _____
 Carrier: FedEx - UPS - DHL - USPS Method: GRD - NDA - ~~2DY~~ - ~~3DY~~ - Other

Product Information:

Product: 7410-~~7110~~-6510-6810-7510 Serial #: AR WL-0008
 9510- Drug Tester 5000
 Description: A - B - Plus - Screener - Demo Printer Ser #: AR -
 Whole Instrument Top ½ Sim Ser #: _____
 Other _____ Probe Ser # DD P -
ACCESSORIES:
 110 V A/C Adapter Regulator Mag Card Rdr# _____
 Printer Paper Printer Ribbon Casio # _____
 Mouthpieces Carrying Case Dry Gas
 Other (Please Specify) _____ Warranty Exp. Date _____

Repair Information:

Test #

Reason for Return: _____

Part Number	Description	Qty	Total Cost
<u>MP Cal 71</u>	<u>Cal</u>	<u>1</u>	<u>N/A</u>
<u>MP Labor</u>	<u>labor</u>	<u>.5</u>	<u>N/A</u>

Repair Notes:

Removed ERROR 112

Cal w/DC and Ops Check

Service Technician: _____

Date: 11/25/09