

Return & Repair Form:

Customer Information:

Company Name: BELMANR NJ B: _____ S: _____
 Date Received: 1-5-09 Date Given to Service: 1-5-09
 Carrier: ~~FedEx~~ - UPS - DHL - USPS Method: GRD - NDA - 2DY - 3DY - Other

Product Information:

Product: 7410 - 7110 - 6510 - 6810 Serial #: AR WCL - 0014
 Description: A - B - Plus - Screener - Demo Printer Ser #: AR -
 Whole Instrument Top ½ Sim Ser #: _____
 Other _____ Probe Ser # DD P -

ACCESSORIES:

110 V A/C Adapter Regulator Mag Card Rdr# _____
 Printer Paper Printer Ribbon Casio # _____
 Mouthpieces Carrying Case Dry Gas
 Other (Please Specify) _____ Warranty Exp. Date _____

Repair Information:

Test #

Reason for Return: will NOT COME READY - BREATH HOSE NOT

Part Number	Description	Qty	Total Cost
<u>MP CAL 71</u>	<u>CAL</u>	<u>1</u>	<u>N/C WARR</u>

Repair Notes:

UNIT WAS IN FOR SERVICE 12-19-08 FOR FAULTY SOLDER JOINT
ON BREATH HOSE. - RESET A/D VALUE TO 1000 - TEMP MOVING
DOWN INTO SPEC. TEMP. - OK NOW
SECURITY KEY WAS WRONG? - RESET

CAL W QC E OPS CHECKS

Service Technician: [Signature]

Date: 01-06-09