

# RETURN AND REPAIR FORM



## Customer Information

B:

S:

Company Name: New Jersey State Police - Bellmawr Station, Bellmawr, NJ

Date Received: 01/27/15

Date given to service: 01/27/15

Carrier:  FedEx  UPS  USPS Shipping Method:  GRD  3DAY  2DAY  
 NDA-PRI  NDA-STD

Product:  6510  6810  
 7110  8610  
 7510  DT5000  
 7410 Upper-half  9510  
 7410 Whole

Serial Number: AR WC-0014

Printer Serial#: AR

Sim Serial#: DD

Probe Serial#: DD

Warranty Expires: January 2015

Description:  A  B  Plus  Demo  Screener  Trade In

## Accessories

110V A/C Adapter  Regulator  Printer Ribbon  Printer Paper  
 Mouthpieces  9510 Stylus  9510 Top Cover  Carrying Case  
 Dry Gas  Other (please specify) \_\_\_\_\_

## Repair Information:

Test#:

| Part Number | Description   | Qty | Total Cost |
|-------------|---------------|-----|------------|
| 4414161     | Calibration   | 1   | NC/W       |
| 4414166     | Labor         | 3   | NC/W       |
| 6808455     | Fuel Cell     | 1   | NC/W       |
| 6808486     | Spacer Plates | 2   | NC/W       |
|             |               |     |            |
|             |               |     |            |
|             |               |     |            |
|             |               |     |            |

Repair Notes: Replaced defective fuel cell, spacer plates and resoldered heater hose wires.

Resoldering heater hose wires resolved Error 75 that unit was giving.

Cleaned printer head due to not printing properly.

Fuel Cell = ARFK-0524

CAL W/QC AND OPS CHECK

Service Technician BC

Date: 01/29/15