

Dräger safety

Return & Repair Form

Customer Information:

CUSTOMER New Jersey State Police Cust # _____
 Date of Return: 4/2/7 Carrier: UPS Method: 3-Day

Product Information:

Product : 7110 7410 6510 6810 Serial # ARWC-0018
Description: A B PLUS SCREENER Printer Serial # _____
Whole Inst. Top Half Simulator Serial # _____
OTHER : _____ Probe Serial # _____
 Demo Unit Returned to stock on: _____
Accessories: (check all that apply):
 110VAC Adapter. Regulator Mag Card Rdr # _____
 Printer Paper Printer Ribbon Casio # _____
 Mouthpieces Carrying Case Dry Gas
 Other (specify): _____
 Warranty Expiration Date: _____

Repair Information:

Test # _____

Reason for Return: Random IR + EC results

Part #	Description	Qty	Total
<u>6809512</u>	<u>Hose barb w/check valve</u>	<u>1</u>	
<u>NPLABOR</u>	<u>LABOR</u>	<u>.5</u>	
<u>NPCAL71</u>	<u>7110 CAL</u>	<u>1</u>	

Repair Notes: Check valve was completely stuck.
Replaced check valve. Full cal ops check
+ Q.C.

Technician: CRD

Date: 4/3/07