

# Dräger safety

## Return & Repair Form

49248

### Customer Information:

CUSTOMER New Jersey ST. Police Cust # \_\_\_\_\_

Date of Return: 4-27-07 Carrier: UPS Method: 3-Day

### Product Information:

Product: 7110 7410 6510 6810 Serial # ARWC-0018

Description: A B PLUS SCREENER Printer Serial # \_\_\_\_\_

Whole Inst. Top Half Simulator Serial # \_\_\_\_\_

OTHER: \_\_\_\_\_ Probe Serial # \_\_\_\_\_

Demo Unit  Returned to stock on: \_\_\_\_\_

### Accessories: (check all that apply):

- 110VAC Adapter       Regulator       Mag Card Rdr # \_\_\_\_\_
- Printer Paper       Printer Ribbon       Casio # \_\_\_\_\_
- Mouthpieces       Carrying Case       Dry Gas
- Other (specify): \_\_\_\_\_

Warranty Expiration Date: \_\_\_\_\_

### Repair Information:

Test # \_\_\_\_\_

Reason for Return: Error 032

Part #	Description	Qty	Total
<u>60810269</u>	<u>EC Motor</u>	<u>1</u>	
<u>MP LABOR</u>	<u>LABOR</u>	<u>.5</u>	

Repair Notes: Confirmed E032, Motor did not turn.  
Replaced EC motor. Verified Accuracy

Technician: CRD

Date: 5/1/07