

RETURN AND REPAIR FORM



Customer Information

B:

S:

Company Name: N.J.S.P.- Woodbine , Woodbine Borough, NJ

Date Received: 05/22/12

Date given to service: 05/22/12

Carrier: FedEx UPS USPS Shipping Method: GRD 3DAY 2DAY
 NDA-PRI NDA-STD

Product: 6510 6810
 7110 8610
 7510 DT5000
 7410 Upper-half 9510
 7410 Whole

Serial Number: AR WC-0018

Printer Serial#: AR

Sim Serial#: DD

Probe Serial#: DD

Warranty Expires: _____

Description: A B Plus Demo Screener Trade In

Accessories

110V A/C Adapter Regulator Printer Ribbon Printer Paper
 Mouthpieces 9510 Stylus 9510 Top Cover Carrying Case
 Dry Gas Other (please specify) _____

Repair Information:

Test#:

Part Number	Description	Qty	Total Cost
MP Cal 71	Calibration	1	NC/W
6808455	Fuel Cell	1	NC/W
8315075	PCB(Mother board)	1	NC/W
MP Labor	Labor	.5	NC/W

Repair Notes: Replaced defective fuel cell with fuel cell number ARDD-1080, Unit was also sent back for corrupt memory, replaced PCB.

CAL W/QC AND OPS CHECK

Service Technician JS

Date: 05-30-12