

RETURN AND REPAIR FORM



Customer Information

B:

S:

Company Name: NJSP Bellmawr

Date Received: 8-30-11 Date given to service: 8-30-11

Carrier: FedEx UPS USPS Shipping Method: GRD 3DAY 2DAY
 NDA-PRI NDA-STD

Product: 6510 6810
 7110 8610
 7510 DT5000
 7410 Upper-half 9510
 7410 Whole

Serial Number: AR WC - 0019
 Printer Serial#: AR
 Sim Serial#: DD
 Probe Serial#: DD

Warranty Expires: _____

Description: A B Plus Demo Screener Trade In

Accessories

110V A/C Adapter Regulator Printer Ribbon Printer Paper
 Mouthpieces 9510 Stylus 9510 Top Cover Carrying Case
 Dry Gas Other (please specify) _____

Repair Information:

Test#: _____

Part Number	Description	Qty	Total Cost
8312820	Pump	1	n/c w
8315075	PCB Board	1	n/c w
8100902	Flow Sensor	1	n/c w
mp cal 71	Calibrate	1	n/c w
mp labor	labor	.5	n/c w

Repair Notes: Unit arrived with a error (009), unit also gave a
blowing not allowed, the pump also made a dragging
noise while purging. The PCB Board and pump were
replaced. also had to replace broken Flow Sensor.
cal w/ Dräger ops check

Service Technician: [Signature]

Date: 8-31-11