

RETURN AND REPAIR FORM



Customer Information

B:

S:

Company Name: N.J. State Police, Hamilton, NJ

Date Received: 09/03/14

Date given to service: 09/03/14

Carrier: FedEx UPS USPS Shipping Method: GRD 3DAY 2DAY
 NDA-PRI NDA-STD

Product: 6510 6810
 7110 8610
 7510 DT5000
 7410 Upper-half 9510
 7410 Whole

Serial Number: AR WC-0019

Printer Serial#: AR

Sim Serial#: DD

Probe Serial#: DD

Warranty Expires: January 2015

Description: A B Plus Demo Screener Trade In

Accessories

110V A/C Adapter Regulator Printer Ribbon Printer Paper
 Mouthpieces 9510 Stylus 9510 Top Cover Carrying Case
 Dry Gas Other (please specify) _____

Repair Information:

Test#:

Part Number	Description	Qty	Total Cost
4414161	Calibration	1	
4414166	Labor	1	
6808455	Fuel Cell	1	
6808486	Spacer Plates	2	

Repair Notes: Replaced defective C156, C167, fuel cell, and spacer plates.

Changing C156 and C167 resolved the issue customer reported having with modem. (line busy)

Fuel Cell = ARFD-1156

CAL W/QC AND OPS CHECK

Service Technician BC

Date: 09/05/14