

RETURN AND REPAIR FORM



Customer Information

B:

S:

Company Name: N.J. SP. Bellmawr, Bellmawr, NJ

Date Received: 08/02/11 Date given to service: _____

Carrier: FedEx UPS USPS Shipping Method: GRD 3DAY 2DAY
 NDA-PRI NDA-STD

Product: 6510 6810
 7110 8610
 7510 DT5000
 7410 Upper-half 9510
 7410 Whole

Serial Number: AR WC-0019

Printer Serial#: AR

Sim Serial#: DD

Probe Serial#: DD

Warranty Expires: _____

Description: A B Plus Demo Screener Trade In

Accessories

110V A/C Adapter Regulator Printer Ribbon Printer Paper
 Mouthpieces 9510 Stylus 9510 Top Cover Carrying Case
 Dry Gas Other (please specify) _____

Repair Information:

Test#:

Part Number	Description	Qty	Total Cost
<u>mp cal 71</u>	<u>calibrate</u>	<u>1</u>	<u>n/c w</u>
<u>12029</u>	<u>power supply</u>	<u>1</u>	<u>n/c w</u>
<u>4412001</u>	<u>modern Jack</u>	<u>1</u>	<u>n/c w</u>
<u>mp labor</u>	<u>labor</u>	<u>.5</u>	<u>n/c w</u>

Repair Notes: Install new power supply to restore power to unit.

memory was also downloaded before calibration.
cal w/QC; ops check

Service Technician: [Signature]

Date: 8-12-11