

# RETURN AND REPAIR FORM



### Customer Information

Company Name: N.J.S.P. Norris, Port Norris, NJ

Date Received: 11/08/11 Date given to service: \_\_\_\_\_

Carrier:  FedEx  UPS  USPS Shipping Method:  GRD  3DAY  2DAY  
 NDA-PRI  NDA-STD

Product:  6510  6810  
 7110  8610  
 7510  DT5000  
 7410 Upper-half  9510  
 7410 Whole

Serial Number: AR WC-0054  
 Printer Serial#: AR  
 Slim Serial#: DD  
 Probe Serial#: DD

Warranty Expires: \_\_\_\_\_

Description:  A  B  Plus  Demo  Screener  Trade In

### Accessories

- 110V A/C Adapter
- Regulator
- Printer Ribbon
- Printer Paper
- Mouthpieces
- 9510 Stylus
- 9510 Top Cover
- Carrying Case
- Dry Gas
- Other (please specify) \_\_\_\_\_

### Repair Information:

Test#: \_\_\_\_\_

Part Number	Description	Qty	Total Cost
<u>MP cel 71</u>	<u>Calibration</u>	<u>1</u>	<u>N/C W</u>
<u>0808455</u>	<u>Fuel Cell</u>	<u>1</u>	<u>N/C W</u>
<u>MP Labor</u>	<u>Labor</u>	<u>.5</u>	<u>N/C W</u>

Repair Notes: Replaced Fuel Cell ARXD-1383, with Fuel Cell ARCH-1002.

cal w/pc, ops check

Service Technician: [Signature] Date: 11-9-11