

Return & Repair Form: 10383878
80490964

Customer Information:

Company Name: NJSP - Burlington B: 53009 S: 85227
 Date Received: 1-4-10 Date Given to Service: _____
 Carrier: FedEx - UPS - DHL - USPS Method: GRD - NDA - 2DY - 3DY - Other

Product Information:

Product: 7410-7110-6510-6810-7510 Serial #: AR WC - 0056
 9510- Drug Tester 500C
 Description: A - B - Plus - Screener - Demo Printer Ser #: AR
 Whole Instrument Top 1/2 Sim Ser #: _____
 Other _____ Probe Ser # DD P
ACCESSORIES:
 110 V A/C Adapter Regulator Mag Card Rdr# _____
 Printer Paper Printer Ribbon Casio # _____
 Mouthpieces Carrying Case Dry Gas
 Other (Please Specify) _____ Warranty Exp. Date _____

Repair Information:

Test #

Reason for Return: PCB

Part Number	Description	Qty	Total Cost
<u>8315075</u>	<u>PCB</u>	<u>1</u>	<u>N/C W/AVR</u>
<u>Q808453</u>	<u>FC</u>	<u>1</u>	<u>N/C W/AVR</u>
<u>MP Cal 71</u>	<u>CAL</u>	<u>1</u>	<u>N/C W/AVR</u>
<u>MP Labor</u>	<u>Labor</u>	<u>.5</u>	<u>N/C W/AVR</u>

Repair Notes:

Install FC # ARMS-0029 / REPLACED PCB
Cal w/DC; Ops Check

Service Technician: AS

Date: 01/06/2010