RETURN AND REPAIR FORM

Dräger

Customer Information

B: S:

Company Name: Garwood P.D., Garwood, NJ

Date Received: 10/05/2016 Date given to service: 10/05/2016

Carrier: ☒ FedEx ☐ UPS ☐ USPS Shipping Method: ☐ GRD ☒ 3DAY ☐ 2DAY ☐ NDA-PRI ☐ NDA-STD

Product:
- ☐ 6510
- ☒ 7110
- ☐ 7510
- ☐ 7410 Upper-half
- ☐ 7410 Whole
- ☐ 6810
- ☐ 8610
- ☐ DT5000
- ☐ 9510

Serial Number: AR WC-0057

Printer Serial#: AR

Sim Serial#: DD

Probe Serial#: DD

Warranty Expires:

Description: ☐ A ☐ B ☐ Plus ☐ Demo ☐ Screener ☐ Trade In

Accessories

- ☐ 110V A/C Adapter
- ☐ Mouthpieces
- ☐ Dry Gas
- ☐ Regulator
- ☐ 9510 Stylus
- ☐ Other (please specify)
- ☐ Printer Ribbon
- ☐ 9510 Top Cover
- ☐ Printer Paper
- ☐ Carrying Case

Repair Information

Test#:

Part Number	Description	Qty	Total Cost
4414161	Calibration	1	
4414166	Labor	.5	
6808455	Fuel Cell	1	
6808486	Spacer Plates	2	

Repair Notes: Replaced defective fuel cell and spacer plates.

Fuel Cell = ARJE-2065

CAL W/QC AND OPS CHECK

Service Technician BC Date: 11/03/2016

sustnf6010\Public\Forms\return and repair form 07_21_2011.pdf