

Return & Repair Form

Customer Information:

CUSTOMER Willingboro Twpsh Cust # 1500-57577 Date of Return: 1/16/2008 Carrier: (FX) UPS DHL USPS Method: GRD (NDA) 2DY 3DY OTHER: STD

S+B

Product Information:

Product: 7110 7410 6510 6810 Serial # AR WC - 0060
Description: A B PLUS SCREENER Printer Serial # AR -
Whole Inst. Top Half Simulator Serial # _____
OTHER: _____ Probe Serial # DD P -

Demo Unit Returned to stock on: _____

Accessories: (check all that apply):

- 110VAC Adapter Regulator Mag Card Rdr # _____
- Printer Paper Printer Ribbon Casio # _____
- Mouthpieces Carrying Case Dry Gas
- Other (specify): _____

Warranty Expiration Date: _____

Repair Information:

Test # _____

Reason for Return: FC RESULTS LOW

Part #	Description	Qty	Total
<u>6808455</u>	<u>FC</u>	<u>1</u>	<u>N/C WARR</u>
<u>MPCH 71</u>	<u>7110 CAB</u>	<u>1</u>	<u>N/C WARR</u>
<u>MPLABOR</u>	<u>LABOR</u>	<u>.5</u>	<u>N/C WARR</u>

Repair Notes: REPLACE FC / CAB w/ QC & OPS CHECKS

Technician: [Signature]

Date: 01-21-08