

# Return & Repair Form:

## Customer Information:

Company Name: South Toms River B: \_\_\_\_\_ S: \_\_\_\_\_

Date Received: 7-17-2009 Date Given to Service: \_\_\_\_\_

Carrier: FedEx - UPS - DHL - USPS Method: GRD - NDA - 2DY 3DY - Other

## Product Information:

Product: 7410 - 7110 - 6510 - 6810 Serial #: AR INC - 0061

Description: A - B - Plus - Screener - Demo Printer Ser #: AR -

Whole Instrument Top 1/2 Sim Ser #: \_\_\_\_\_

Other \_\_\_\_\_ Probe Ser # DD \_\_\_\_\_ P -

## ACCESSORIES:

110 V A/C Adapter  Regulator  Mag Card Rdr# \_\_\_\_\_

Printer Paper  Printer Ribbon  Casio # \_\_\_\_\_

Mouthpieces  Carrying Case  Dry Gas

Other (Please Specify) \_\_\_\_\_ Warranty Exp. Date \_\_\_\_\_

## Repair Information:

FC, PCB Test #

Reason for Return: \_\_\_\_\_

Part Number	Description	Qty	Total Cost
<u>8315075</u>	<u>PCB</u>	<u>1</u>	<u>N/C W/AVY</u>
<u>6808455</u>	<u>FC</u>	<u>1</u>	<u>N/C W/AVY</u>
<u>688486</u>	<u>Plate</u>	<u>1</u>	<u>N/C W/AVY</u>
<u>MP Cal 71</u>	<u>Cal</u>	<u>1</u>	<u>N/C W/AVY</u>
<u>MP Cal 60</u>	<u>Cal 60</u>	<u>1.0</u>	<u>N/C W/AVY</u>

Repair Notes:

Replace FC, PCB

Cal w/Qc j Ops Check

Service Technician: [Signature]

Date: 7/22/2009