

Return & Repair Form:

Customer Information:

Company Name: Pemberton Township S: _____
 Date Received: 8-31-2010 Date Given to Service: _____
 Carrier: FedEx - UPS - DHL - USPS Method: GRD - NDA - 2DY - 3DY - Other

Product Information:

Product: 7410-7110-6510-6810-7510
 9510-Drug Tester 5000 Serial #: ARWC-0063
 Description: A - B - Plus - Screener - Demo Printer Ser #: AR
 Whole Instrument Top ½ Sim Ser #: _____
 Other _____ Probe Ser # DD _____ P _____

ACCESSORIES:

110 V A/C Adapter Regulator Mag Card Rdr# _____
 Printer Paper Printer Ribbon Casio # _____
 Mouthpieces Carrying Case Dry Gas
 Other (Please Specify) _____ Warranty Exp. Date _____

Repair Information:

Test #

Reason for Return:

FC / motor

Part Number	Description	Qty	Total Cost
<u>6808455</u>	<u>FC</u>	<u>1</u>	<u>N/C W</u>
<u>6808374</u>	<u>MOTOR</u>	<u>1</u>	<u>N/C W</u>
<u>6808486</u>	<u>PLATE</u>	<u>1</u>	<u>N/C W</u>
<u>MP Cal 71</u>	<u>Cal</u>	<u>1</u>	<u>N/C W</u>
<u>MP Labor</u>	<u>Labor</u>	<u>.5</u>	<u>N/C W</u>

Repair Notes:

NEW FC # AR13E-1108

New Cap tube

Cal w/DC & OPS Check

Service Technician: _____

[Signature]

Date: _____

9-1-2010