

# RETURN AND REPAIR FORM



## Customer Information

B:

S:

Company Name: Pemberton Twp. P.D., Pemberton, NJ

Date Received: 10/10/13

Date given to service: 10/10/13

Carrier:  FedEx  UPS  USPS

Shipping Method:  GRD  3DAY  2DAY  
 NDA-PRI  NDA-STD

Product:  6510  6810  
 7110  8610  
 7510  DT5000  
 7410 Upper-half  9510  
 7410 Whole

Serial Number: AR WC-0063

Printer Serial#: AR

Sim Serial#: DD

Probe Serial#: DD

Warranty Expires: October 2016

Description:  A  B  Plus  Demo  Screener  Trade In

## Accessories

110V A/C Adapter  Regulator  Printer Ribbon  Printer Paper  
 Mouthpieces  9510 Stylus  9510 Top Cover  Carrying Case  
 Dry Gas  Other (please specify) \_\_\_\_\_

## Repair Information:

Test#:

Part Number	Description	Qty	Total Cost
MP Cal 71	Calibration	1	NC/W
MP Labor	Labor	2	NC/W

Repair Notes: Replaced defective C156 and C167. Complaint was when connecting with modem it would give a line busy signal and issue was resolved with changing the two capacitors.

CAL W/QC AND OPS CHECK

Service Technician BC

Date: 10/15/13