

RETURN AND REPAIR FORM



Customer Information

B: _____ S: _____

Company Name: N.J. State Police-Woodstown , Pilesgrove, NJ

Date Received: 02-28-12

Date given to service: 02-28-12

Carrier: FedEx UPS USPS Shipping Method: GRD 3DAY 2DAY
 NDA-PRI NDA-STD

Product: 6510 6810
 7110 8610
 7510 DT5000
 7410 Upper-half 9510
 7410 Whole

Serial Number: AR WC-0067

Printer Serial#: AR

Sim Serial#: DD

Probe Serial#: DD

Warranty Expires: _____

Description: A B Plus Demo Screener Trade In

Accessories

110V A/C Adapter Regulator Printer Ribbon Printer Paper
 Mouthpieces 9510 Stylus 9510 Top Cover Carrying Case
 Dry Gas Other (please specify) _____

Repair Information:

Test#: _____

Part Number	Description	Qty	Total Cost
8315075	PCB Board	1	N/C W
MP CAL 71	Calibration	1	N/C W
MP Labor	Labor	.5	N/C W

Repair Notes: While the customer was performing a part two linearity test the instrument displayed and reported a different concentration then the 0.080 that was entered. CAL / W QC AND OPS CHECK

Service Technician: JS

Date: 04-24-12