

Dräger safety

Return & Repair Form

Customer Information:

CUSTOMER Plainfield Police Cust # 57486/64482
 Date of Return: 11-6-06 Carrier: FP Method: PRI

Product Information:

Product: 7110 7410 6510 6810 Serial # ARWC-0069
 Description: A B PLUS SCREENER _ Printer Serial # _____
 Whole Inst. Top Half Simulator Serial # _____
 OTHER: _____ Probe Serial # _____
 Demo Unit Returned to stock on: _____
 Accessories: (check all that apply):
 110VAC Adapter Regulator Mag Card Rdr # _____
 Printer Paper Printer Ribbon Casio # _____
 Mouthpieces Carrying Case Dry Gas
 Other (specify): _____
 Warranty Expiration Date: _____

Repair Information:

Test # _____

Reason for Return: EC + IR dropping out

Part #	Description	Qty	Total
6808455	Fuel Cell	1	
6809512	Cuvette inlet w/check valve	1	
MPLABOR	LABOR	1.0	
IMPCAL71	7110 CAL	1	

Repair Notes: Replaced questionable fuel cell and questionable check valve in cuvette. Full cal ops check + QC.

Technician: CRD

Date: 11/8/06