

# Return & Repair Form:

## Customer Information:

Company Name: Trenton PD B: \_\_\_\_\_ S: \_\_\_\_\_

Date Received: 05-18-2009 Date Given to Service: 05-18-2009

Carrier: FedEx - UPS - DHL - USPS Method: GRD - NDA - 2DY - 3DY Other

## Product Information:

Product: 7410 - 7110 - 6510 - 6810

Serial #: AR WE-0002

Description: A - B - Plus - Screener - Demo

Printer Ser #: AR -

Whole Instrument

Top 1/2

Sim Ser #: \_\_\_\_\_

Other \_\_\_\_\_

Probe Ser # DD P -

## ACCESSORIES:

110 V A/C Adapter

Regulator

Mag Card Rdr# \_\_\_\_\_

Printer Paper

Printer Ribbon

Casio # \_\_\_\_\_

Mouthpieces

Carrying Case

Dry Gas

Other (Please Specify) \_\_\_\_\_

Warranty Exp. Date \_\_\_\_\_

## Repair Information:

FC

Test #

Reason for Return:

Part Number	Description	Qty	Total Cost
<u>6808486</u>	<u>Plate</u>	<u>1</u>	<u>N/C Warr</u>
<u>6808455</u>	<u>FC</u>	<u>1</u>	<u>N/C Warr</u>
<u>MP Cal 71</u>	<u>Cal</u>	<u>1</u>	<u>N/C Warr</u>
<u>MP Labor</u>	<u>Labor</u>	<u>.5</u>	<u>N/C Warr</u>

Repair Notes:

Replace FC

Cal w/DC & Ops Check

Service Technician: [Signature]

Date: 05-19-2009