

SENT PRIORITY ON SATURDAY DELIVERY WHY?

# Dräger safety

Internal Use Only

→ 1500-58755

## Return & Repair Form

BTS J

### Customer Information:

CUSTOMER Washington Top. Police Dept. Cust # 1500-58756  
 Date of Return: 02/15/2008 Carrier:  **UPS**  **DHL**  **USPS** Method:  **GRD**  **NDA**  **2DY**  **3DY**  
 OTHER: \_\_\_\_\_

### Product Information:

Product : 7110 7410 6510 6810 Serial # AR WE - 0004  
 Description: A B PLUS SCREENER Printer Serial # AR -  
 Whole Inst. Top Half Simulator Serial # \_\_\_\_\_  
 OTHER : \_\_\_\_\_ Probe Serial # DD P -  
 Demo Unit  Returned to stock on: \_\_\_\_\_  
 Accessories: (check all that apply):  
 110VAC Adapter  Regulator  Mag Card Rdr # \_\_\_\_\_  
 Printer Paper  Printer Ribbon  Casio # \_\_\_\_\_  
 Mouthpieces  Carrying Case  Dry Gas  
 Other (specify): \_\_\_\_\_  
 Warranty Expiration Date: \_\_\_\_\_

### Repair Information:

Test # \_\_\_\_\_

Reason for Return: \_\_\_\_\_

Part #	Description	Qty	Total
<u>6808455</u>	<u>FC</u>	<u>1</u>	<u>N/C WARR</u>
<u>MPCAL 71</u>	<u>7110 CAL</u>	<u>1</u>	<u>N/C WARR</u>
<u>MPLABOR</u>	<u>LABOR</u>	<u>.5</u>	<u>N/C WARR</u>

Repair Notes: REPLACE FC / CAL w/ QC & OPS CHECKS

Technician: [Signature]

Date: 02-21-08