

Return & Repair Form:

Customer Information:

Company Name: Lanvale Springs NJ B: _____ S: _____
 Date Received: 1 Jan 11 Date Given to Service: 1 Jun 11
 Carrier: FedEx - UPS - DHL - USPS Method: GRD - NDA - 2DY - 3DY - Other

Product Information:

Product: 7410-7110-6510-6810-7510
 9510-Drug Tester 5000
 Description: A - B - Plus - Screener - Demo
 Serial #: ARWE - 0005
 Printer Ser #: AR -
 Sim Ser #: _____
 Probe Ser # DD P -
 Whole Instrument Top 1/2
 Other _____
ACCESSORIES:
 110 V A/C Adapter Regulator Mag Card Rdr# _____
 Printer Paper Printer Ribbon Casio # _____
 Mouthpieces Carrying Case Dry Gas
 Other (Please Specify) _____ Warranty Exp. Date _____

Repair Information:

Test #

Reason for Return: _____

Part Number	Description	Qty	Total Cost
<u>6808455</u>	<u>FC</u>	<u>1</u>	<u>n/c w</u>
<u>mp cal 71</u>	<u>Cal</u>	<u>1</u>	<u>n/c w</u>
<u>mp labor</u>	<u>Labor</u>	<u>.5</u>	<u>n/c w</u>

Repair Notes:

FC # ARAN-0942 replaced, because of a error (31), FC sent back to Germany.

Cal w/DC; oys check

Service Technician: [Signature]

Date: 6-7-11