

RETURN AND REPAIR FORM



Customer Information

B:

S:

Company Name: N.J.S.P. Sussex, Frankford NJ

Date Received: 05/15/12

Date given to service: 05/15/12

Carrier: FedEx UPS USPS Shipping Method: GRD 3DAY 2DAY
 NDA-PRI NDA-STD

Product: 6510 6810
 7110 8610
 7510 DT5000
 7410 Upper-half 9510
 7410 Whole

Serial Number: AR WE-0008

Printer Serial#: AR

Sim Serial#: DD

Probe Serial#: DD

Warranty Expires: _____

Description: A B Plus Demo Screener Trade In

Accessories

110V A/C Adapter Regulator Printer Ribbon Printer Paper
 Mouthpieces 9510 Stylus 9510 Top Cover Carrying Case
 Dry Gas Other (please specify) _____

Repair Information

Test#:

| Part Number | Description | Qty | Total Cost |
|-------------|-------------|-----|------------|
| MP CAL 71 | Calibration | 1 | N/C W |
| 6808455 | Fuel cell | 1 | N/CW |
| MP Labor | Labor | .5 | N/C W |

Repair Notes: Replaced defective fuel cell with fuel cell ARCM-1767.

CAL W/QC AND OPS CHECK

Service Technician JS

Date: 05-21-12