RETURN AND REPAIR FORM

Dräger

Customer Information

B: S:

Company Name: Springfield TWP P.D., Jobstown, NJ

Date Received: 08/03/11 Date given to service:

Carrier: [X] FedEx [ ] UPS [ ] USPS

Shipping Method: [ ] GRD [X] 3DAY [ ] 2DAY [ ] NDA-PRI [ ] NDA-STD

Product: [ ] 6510 [ ] 6810 [X] 7110 [ ] 8610 [ ] 7510 [ ] DT5000 [ ] 7410 Upper-half [ ] 9510 [ ] 7410 Whole

Serial Number: AR WE-0012

Printer Serial#: AR

Sim Serial#: DD

Probe Serial#: DD

Warranty Expires:

Description: [ ] A [ ] B [ ] Plus [ ] Demo [ ] Screener [ ] Trade In

Accessories

[ ] 110V A/C Adapter [ ] Regulator [ ] Printer Ribbon [ ] Printer Paper
[ ] Mouthpieces [ ] 9510 Stylus [ ] 9510 Top Cover [ ] Carrying Case
[ ] Dry Gas [ ] Other (please specify)

Repair Information

Test#:

Part Number	Description	Qty	Total Cost
Misc Part	7110 Reset button	1	N/C W
mp Cal 71	Cal	1	N/C W
mp labor	labor	.5	N/C W

Repair Notes: Unit was inspected and tech confirmed 7110 would not reset by pushing orange button, replaced reset button and recalibrated unit.

Service Technician: JS Date: 8-8-11

sustxfs010\Public\Forms\return and repair form 07_21_2011.pdf