

Dräger safety

Return & Repair Form

Customer Information:

CUSTOMER Beach Haven Police Dep. Cust # _____
 Date of Return: 6-4-7 Carrier: FedEx Method: 2-Day

Product Information:

Product: 7110 7410 6510 6810 Serial # ARWE-0013
Description: A B PLUS SCREENER Printer Serial # _____
Whole Inst. Top Half Simulator Serial # _____
OTHER: _____ Probe Serial # _____
 Demo Unit Returned to stock on: _____ *** NOT PACKAGED GOOD.**
Accessories: (check all that apply):
 110VAC Adapter Regulator Mag Card Rdr # _____
 Printer Paper Printer Ribbon Casio # _____
 Mouthpieces Carrying Case Dry Gas
 Other (specify): _____
 Warranty Expiration Date: _____

Repair Information:

Test # _____

Reason for Return: FL Problem

Part #	Description	Qty	Total
<u>0808455</u>	<u>Fuel Cell</u>	<u>1</u>	
<u>MPLABOR</u>	<u>LABOR</u>	<u>.5</u>	
<u>MPCAL71</u>	<u>7110 CAL</u>	<u>1</u>	

Repair Notes: Replaced Questionable Flc. Full cal ops check + Q.C.

Technician: CRO

Date: 6/5/07