

Return & Repair Form:

Customer Information:

Company Name: Winfield Twp B: 58353
 Date Received: 6-8-10 Date Given to Service: _____
 Carrier: FedEx - UPS - DHL - USPS Method: GRD - NDA - 2DY - 3DY - Other

Product Information:

Product: 7410-7110-6510-6810-7510 Serial #: AR WE - 0019
 9510- Drug Tester 5000
 Description: A - B - Plus - Screener - Demo Printer Ser #: AR -
 Whole Instrument Top 1/2 Sim Ser #: _____
 Other _____ Probe Ser # DD P -
ACCESSORIES:
 110 V A/C Adapter Regulator Mag Card Rdr# _____
 Printer Paper Printer Ribbon Casio # _____
 Mouthpieces Carrying Case Dry Gas
 Other (Please Specify) _____ Warranty Exp. Date JULY 2010

Repair Information:

Test #

Reason for Return: FC/motor

Part Number	Description	Qty	Total Cost
<u>6808374</u>	<u>motor</u>	<u>1</u>	<u>N/C W</u>
<u>6808486</u>	<u>Plate</u>	<u>1</u>	<u>N/C W</u>
<u>6808455</u>	<u>FC</u>	<u>1</u>	<u>N/C W</u>
<u>MP cal 71</u>	<u>cal</u>	<u>1</u>	<u>N/C W</u>
<u>MP labor</u>	<u>labor</u>	<u>1</u>	<u>N/C W</u>

Repair Notes:

FC# AR130-2565 unit came in with a ERRO 32 bad motor installed new motor.

Cal w/QC & ops check

Service Technician: [Signature]

Date: 6-9-2010