

RETURN AND REPAIR FORM

Customer Information

B:

S:

Company Name: Seaside Park P.D. Seaside Park NJ.

Date Received: 02/06/12 Date given to service: 02/06/12

Carrier: FedEx UPS USPS Shipping Method: GRD 3DAY 2DAY
 NDA-PRI NDA-STD

Product: 6510 6810 7110 7510 7410 Upper-half 7410 Whole 8610 8610 DT5000 9510

Serial Number: AR WE - 0020
 Printer Serial#: AR
 Sim Serial#: DD
 Probe Serial#: DD

Warranty Expires: _____

Description: A B Plus Demo Screener Trade In

Accessories

110V A/C Adapter Regulator Printer Ribbon Printer Paper
 Mouthpieces 9510 Stylus 9510 Top Cover Carrying Case
 Dry Gas Other (please specify) _____

Repair Information:

Test#:

Part Number	Description	Qty	Total Cost
6708/5			
6808/455	Fuel Cell	1	n/c w
mp cel 71	Calibration	1	n/c w
mp labor	Labor	.5	n/c w

Repair Notes: Replaced fuel cell # ARAN-1049, because of leakage
New fuel cell # ARCK-1041.

eq w/ Dr ops check

Service Technician: ps Date: 2/13/2012