

## Return & Repair Form

### Customer Information:

CUSTOMER NJ State Police Cust # \_\_\_\_\_  
Date of Return: 3/19/07 Carrier: Fenex Method: priority

### Product Information:

Product : 7110 7410 6510 6810 Serial # ARWE-0021  
Description: A B PLUS SCREENER Printer Serial # \_\_\_\_\_  
**Whole Inst. Top Half** Simulator Serial # \_\_\_\_\_  
OTHER : \_\_\_\_\_ Probe Serial # \_\_\_\_\_

Demo Unit  Returned to stock on: \_\_\_\_\_

### Accessories: (check all that apply):

- 110VAC Adapter       Regulator       Mag Card Rdr # \_\_\_\_\_  
 Printer Paper       Printer Ribbon       Casio # \_\_\_\_\_  
 Mouthpieces       Carrying Case       Dry Gas  
 Other (specify): \_\_\_\_\_

Warranty Expiration Date: \_\_\_\_\_

### Repair Information:

Test # \_\_\_\_\_

Reason for Return: repair ? IR + EC Problem

Part #	Description	Qty	Total
<u>6808455</u>	<u>Fuel Cell</u>	<u>1</u>	
<u>MPLABOR</u>	<u>LABOR</u>	<u>.5</u>	
<u>MPCAL71</u>	<u>7110 CAL</u>	<u>1</u>	

Repair Notes: Replaced questionable FLC. Could not confirm check valve problem. Full cal ops check + Q.C.

Technician: CRD

Date: 3/21/07