

Return & Repair Form:

Customer Information:

Company Name: New Jersey - Washington B. Station S: _____
 Date Received: 3/3/2011 Date Given to Service: 3/3/2011
 Carrier: FedEx - UPS - DHL - USPS Method: GRD - NDA - 2DY - 3DY - Other

Product Information:

Product: ~~7410-7110-8510-6810-7510~~ 9510-Drug Tester 5000 Serial #: ARWE-0022
 Description: A - B - Plus - Screener - Demo Printer Ser #: AR -
 Whole Instrument Top ½ Sim Ser #: _____
 Other _____ Probe Ser # DD P -

ACCESSORIES:

110 V A/C Adapter Regulator Mag Card Rdr# _____
 Printer Paper Printer Ribbon Casio # _____
 Mouthpieces Carrying Case Dry Gas
 Other (Please Specify) _____ Warranty Exp. Date _____

Repair Information:

Test #

Reason for Return: PCB

Part Number	Description	Qty	Total Cost
<u>8315075</u>	<u>PCB</u>	<u>1</u>	<u>w/cw</u>
<u>MPCAL 91</u>	<u>Cal</u>	<u>1</u>	<u>w/cw</u>
<u>mp labor</u>	<u>labor</u>	<u>.5</u>	<u>w/cw</u>

Repair Notes:

Unit would not come out of standby mode, installed a new PCB.
Cal w/PC & ops check

Service Technician: [Signature]

Date: 3/3/2011