

Return & Repair Form:

**Customer Information:**

Company Name: Prince for TWP B: 57964 S: \_\_\_\_\_  
 Date Received: 10-1-2010 Date Given to Service: \_\_\_\_\_  
 Carrier: FedEx - UPS - DHL - USPS Method: GRD - NDA - 2DY - 3DY - Other

**Product Information:**

Product: 7410-7110-6510-6810-7510 Serial #: AR WF-0224  
 9510- Drug Tester 5000  
 Description: A - B - Plus - Screener - Demo Printer Ser #: AR -  
 Whole Instrument Top 1/2 Sim Ser #: \_\_\_\_\_  
 Other \_\_\_\_\_ Probe Ser # DD P -

**ACCESSORIES:**

110 V A/C Adapter  Regulator  Mag Card Rdr# \_\_\_\_\_  
 Printer Paper  Printer Ribbon  Casio # \_\_\_\_\_  
 Mouthpieces  Carrying Case  Dry Gas  
 Other (Please Specify) \_\_\_\_\_ Warranty Exp. Date OCT. 2013

**Repair Information:**

Test #

Reason for Return: PCB

Part Number	Description	Qty	Total Cost
<u>8315075</u>	<u>PCB</u>	<u>1</u>	<u>n/cw</u>
<u>MT Cal 71</u>	<u>Cal</u>	<u>1</u>	<u>n/cw</u>
<u>MP Labo</u>	<u>Labo</u>	<u>15</u>	<u>n/cw</u>

Repair Notes:

Install new PCB.  
Cal w/QC & ops check

Service Technician: [Signature]

Date: 10-5-2010