

Dräger safety

Return & Repair Form

Please return
standard
overnight

Customer Information:

CUSTOMER RARITAN Twp. Police Dept. a

Date of Return: 3/30/07 Carrier: FEDEX Method: Curtis

Product Information:

Product : 7110 7410 6510 6810

Serial # ARWE-0028

Description: A B PLUS SCREENER

Printer Serial # _____

Whole Inst. Top Half

Simulator Serial # _____

OTHER : _____

Probe Serial # _____

Demo Unit Returned to stock on: _____

Accessories: (check all that apply):

110VAC Adapter Regulator Mag Card Rdr # _____

Printer Paper Printer Ribbon Casio # _____

Mouthpieces Carrying Case Dry Gas

Other (specify): _____

Warranty Expiration Date: _____

Repair Information:

Test # _____

Reason for Return: Error 032

Part #	Description	Qty	Total
<u>MPCAL71</u>	<u>7110 CAL</u>	<u>1</u>	

Repair Notes: Could not confirm E032. Full cal
ops check + Q.C.

Technician: CRD

Date: 4/3/07