

Return & Repair Form

Customer Information:

CUSTOMER Hunterdon County Township of Warren Kelck Dep # 58355
 Date of Return: 8/28/07 Carrier: UP Fedex Method: Priority Overnight

Product Information:

Product: (7110) 7410 6510 6810 Serial # ARWR-0028
 Description: A B PLUS SCREENER Printer Serial # _____
 Whole Inst. Top Half Simulator Serial # _____
 OTHER: _____ Probe Serial # _____

Demo Unit Returned to stock on: _____

Accessories: (check all that apply):

- 110VAC Adapter
- Regulator
- Mag Card Rdr # _____
- Printer Paper
- Printer Ribbon
- Casio # _____
- Mouthpieces
- Carrying Case
- Dry Gas
- Other (specify): _____

Warranty Expiration Date: _____

Repair Information:

Test # _____

Reason for Return: ERROR 32

Part #	Description	Qty	Total
<u>6810269</u>	<u>MOTOR</u>	<u>1</u>	<u>WARRANTY</u>
<u>6807819</u>	<u>SWITCH</u>	<u>1</u>	<u>WARRANTY</u>
<u>6808455</u>	<u>FC</u>	<u>1</u>	<u>WARRANTY</u>

Repair Notes: REPLACED MOTOR & SWITCH (SECOND TIME IN FOR SAME ERROR) FC-WC / SMALL AMOUNT OF ACID LEAK AROUND LEADS - REPLACE CELL AS PM / NEW CELL = 4D-1574 / CAL + QC & OPS CHECKS

Technician: [Signature]

Date: 8-29-07