

Return & Repair Form

Customer Information:

CUSTOMER New Jersey State Police Dept 482481
 Cust # 574405
 Date of Return: 9/27/2007 Carrier: UPS DHL USPS Method: GRD NDA 2DY 3DY
 OTHER: STD ON

Product Information:

Product : 7110 7410 6510 6810 Serial # AR WE - 0029
 Description: A B PLUS SCREENER _ Printer Serial # AR -
 Whole Inst. Top Half Simulator Serial # _____
 OTHER : _____ Probe Serial # DD P -
 Demo Unit Returned to stock on: _____
 Accessories: (check all that apply):
 110VAC Adapter Regulator Mag Card Rdr # _____
 Printer Paper Printer Ribbon Casio # _____
 Mouthpieces Carrying Case Dry Gas
 Other (specify): _____
 Warranty Expiration Date: _____


Repair Information:

Test # _____

Reason for Return: FC PROBLEM

Part #	Description	Qty	Total
<u>6808455</u>	<u>FC</u>	<u>1</u>	<u>WARR</u>
<u>MPCAL7H</u>	<u>7110CA</u>	<u>1</u>	<u>WARR</u>
<u>MPLLABOR</u>	<u>LABOR</u>	<u>.5</u>	<u>WARR</u>

Repair Notes: REPLACED FC / CAL, QC, OPS CHECKS

Technician: 

Date: 09-28-07