

# Dräger safety

## Return & Repair Form

### Customer Information:

CUSTOMER New Jersey State Police - Bordentown Cust # 74687  
 Date of Return: 03 118 12008 Carrier:  UPS  DHL  USPS Method:  GRD  NDA  2DY  3DY  
 OTHER: \_\_\_\_\_

### Product Information:

Product: 7110 7410 6510 6810 Serial # AR WE - 0031  
 Description: A B PLUS SCREENER Printer Serial # AR - \_\_\_\_\_  
 Whole Inst. Top Half Simulator Serial # \_\_\_\_\_  
 OTHER: \_\_\_\_\_ Probe Serial # DD P - \_\_\_\_\_  
 Demo Unit  Returned to stock on: \_\_\_\_\_  
 Accessories: (check all that apply):  
 110VAC Adapter  Regulator  Mag Card Rdr # \_\_\_\_\_  
 Printer Paper  Printer Ribbon  Casio # \_\_\_\_\_  
 Mouthpieces  Carrying Case  Dry Gas  
 Other (specify): \_\_\_\_\_  
 Warranty Expiration Date: \_\_\_\_\_


### Repair Information:

Test # \_\_\_\_\_

Reason for Return: \_\_\_\_\_

Part #	Description	Qty	Total
<u>6808455</u>	<u>FC</u>	<u>1</u>	<u>N/C WARR</u>
<u>MPCR 71</u>	<u>CAL</u>	<u>1</u>	<u>N/C WARR</u>
<u>MPLABOR</u>	<u>LABOR</u>	<u>.5</u>	<u>N/C WARR</u>

Repair Notes: REPLACE FC / CAL W/ QC & OPS CHECKS  
 \_\_\_\_\_  
 \_\_\_\_\_  
 \_\_\_\_\_

Technician: 

Date: 03-20-08