

**Return & Repair Form:**

**Customer Information:**

Company Name: New Jersey State <sup>Police</sup> B: \_\_\_\_\_ S: \_\_\_\_\_  
 Date Received: 12-8-09 Date Given to Service: \_\_\_\_\_  
 Carrier: FedEx - UPS - DHL - USPS Method: GRD - NDA - 2DY - 3DY - Other

**Product Information:**

Product: 7410-7110-6510-6810-7510 Serial #: AR WE - 0031  
 9510- Drug Tester 5000  
 Description: A - B - Plus - Screener - Demo Printer Ser #: AR -  
 Whole Instrument Top 1/2 Sim Ser #: \_\_\_\_\_  
 Other \_\_\_\_\_ Probe Ser # DD P -  
**ACCESSORIES:**  
 110 V A/C Adapter  Regulator  Mag Card Rdr# \_\_\_\_\_  
 Printer Paper  Printer Ribbon  Casio # \_\_\_\_\_  
 Mouthpieces  Carrying Case  Dry Gas  
 Other (Please Specify) \_\_\_\_\_ Warranty Exp. Date \_\_\_\_\_

**Repair Information:**

Test #

Reason for Return:

Part Number	Description	Qty	Total Cost
<u>8315075</u>	<u>PCB</u>	<u>1</u>	<u>w/c warr</u>
<u>MP cal 71</u>	<u>cal</u>	<u>1</u>	<u>w/c warr</u>
<u>MP labor</u>	<u>labor</u>	<u>.5</u>	<u>w/c warr</u>

Repair Notes:

Install New PCB  
cal w/ 20 / 20s check

Service Technician: JS

Date: 12/21/09