

Return & Repair Form:

Customer Information:

Company Name: NEW Jersey PD-HopeB S: _____
 Date Received: 2/18/2011 Date Given to Service: _____
 Carrier: FedEx - UPS - DHL - USPS Method: GRD - NDA - 2DY 3DY - Other

Product Information:

Product: 7410-7110-6510-6810-7510 Serial #: ARWE-0032
 9510- Drug Tester 5000
 Description: A - B - Plus - Screener - Demo Printer Ser #: AR
 Whole Instrument Top ½ Sim Ser #: _____
 Other _____ Probe Ser # DD P

ACCESSORIES:

110 V A/C Adapter Regulator Mag Card Rdr# _____
 Printer Paper Printer Ribbon Casio # _____
 Mouthpieces Carrying Case Dry Gas
 Other (Please Specify) _____ Warranty Exp. Date _____

Repair Information:

Test #

Reason for Return: _____

Part Number	Description	Qty	Total Cost
<u>6807819</u>	<u>Switch</u>	<u>1</u>	<u>n/c w</u>
<u>12043</u>	<u>Tube Kit</u>	<u>1</u>	
<u>mp cal 71</u>	<u>Cal</u>	<u>1</u>	<u>n/c w</u>
<u>mp labor</u>	<u>labor</u>	<u>.5</u>	<u>n/c w</u>

Repair Notes:

ICT # 16-0570 replaced with ICT# BL-1228

cal w/OC ; ops check

Service Technician: [Signature]

Date: 2/21/2011