

# RETURN AND REPAIR FORM



## Customer Information

B:

S:

Company Name: NJSP - Totowa Sub Station, Totowa, NJ

Date Received: 08/27/2016 Date given to service: 08/27/2016

Carrier:  FedEx  UPS  USPS Shipping Method:  GRD  3DAY  2DAY  
 NDA-PRI  NDA-STD

Product:  6510  6810  
 7110  8610  
 7510  DT5000  
 7410 Upper-half  9510  
 7410 Whole

Serial Number: AR WE-0034  
 Printer Serial#: AR  
 Sim Serial#: DD  
 Probe Serial#: DD

Warranty Expires: December 2016

Description:  A  B  Plus  Demo  Screener  Trade In

## Accessories

110V A/C Adapter  Regulator  Printer Ribbon  Printer Paper  
 Mouthpieces  9510 Stylus  9510 Top Cover  Carrying Case  
 Dry Gas  Other (please specify) \_\_\_\_\_

## Repair Information:

Test#:

| Part Number | Description   | Qty | Total Cost |
|-------------|---------------|-----|------------|
| 4414161     | Calibration   | 1   | NC/W       |
| 4414166     | Labor         | 1   | NC/W       |
| 6808455     | Fuel Cell     | 1   | NC/W       |
| 6808374     | Geared Engine | 1   | NC/W       |
|             |               |     |            |
|             |               |     |            |
|             |               |     |            |

Repair Notes: Replaced defective fuel cell and geared engine.  
 \_\_\_\_\_  
 \_\_\_\_\_  
 \_\_\_\_\_

Fuel Cell = ARJC-2220  
 CAL W/QC AND OPS CHECK

Service Technician BC Date: 09/08/2016