

RETURN AND REPAIR FORM



Customer Information

B:

S:

Company Name: NJSP Alcohol/Drug Testing Unit, Trenton, NJ

Date Received: 12/22/2017 Date given to service: 12/22/2017

Carrier: FedEx UPS USPS Shipping Method: GRD 3DAY 2DAY
 NDA-PRI NDA-STD

Product: 6510 6810
 7110 8610
 7510 DT5000
 7410 Upper-half 9510
 7410 Whole

Serial Number: AR WE-0034
 Printer Serial#: AR
 Sim Serial#: DD
 Probe Serial#: DD

Warranty Expires: _____

Description: A B Plus Demo Screener Trade In

Accessories

110V A/C Adapter Regulator Printer Ribbon Printer Paper
 Mouthpieces 9510 Stylus 9510 Top Cover Carrying Case
 Dry Gas Other (please specify) _____

Repair Information:

Test#:

Part Number	Description	Qty	Total Cost
4414161	Calibration	1	
4414166	Labor	1	
6808455	Fuel Cell	1	
6809514	Breath Hose	1	

Repair Notes: Replaced defective fuel cell and breath hose.

Fuel Cell = ARKL-0986
 CAL W/QC AND OPS CHECK

Service Technician BC Date: 01/18/2018