

## Return & Repair Form

### Customer Information:

CUSTOMER New Jersey State Police Cust # \_\_\_\_\_  
 Date of Return: 03 11 2008 Carrier:  UPS  DHL  USPS Method:  GRD  NDA  2D  
 OTHER: Standard

### Product Information:

Product : 7110 7410 6510 6810 Serial # AR WE - 0038  
 Description: A B PLUS SCREENER Printer Serial # AR -  
 Whole Inst. Top Half Simulator Serial # \_\_\_\_\_  
 OTHER : \_\_\_\_\_ Probe Serial # DD P -

Demo Unit  Returned to stock on: \_\_\_\_\_

### Accessories: (check all that apply):

- 110VAC Adapter  Regulator  Mag Card Rdr # \_\_\_\_\_  
 Printer Paper  Printer Ribbon  Casio # \_\_\_\_\_  
 Mouthpieces  Carrying Case  Dry Gas  
 Other (specify): \_\_\_\_\_

Warranty Expiration Date: \_\_\_\_\_

### Repair Information:

Test # \_\_\_\_\_

Reason for Return: \_\_\_\_\_

Part #	Description	Qty	Total
<u>6808455</u>	<u>FC</u>	<u>1</u>	<u>N/CUT</u>
<u>8312835</u>	<u>PRINT HEAD</u>	<u>1</u>	<u>N/CUT</u>
<u>MP CAL 71</u>	<u>7110 CAL</u>	<u>1</u>	<u>N/CUT</u>
<u>MP LABOR</u>	<u>LABOR</u>	<u>1.0</u>	<u>N/CUT</u>

Repair Notes: REPLACE FC - R155 BLOWN ON MOTHER PCB (IN IN  
PRINTER CIRCUIT) REPLACED - PRINT HEAD REQD REPLACEMENT  
CAL W/ QC & OPS CHECKS

Technician: [Signature]

Date: 03-11-2008