

# Return & Repair Form:

## Customer Information:

Company Name: New Jersey State Police B: \_\_\_\_\_ S: \_\_\_\_\_  
 Date Received: 1-18-11 Date Given to Service: \_\_\_\_\_  
 Carrier: FedEx - UPS - DHL - USPS Method: GRD - NDA - 2DY - 3DY - Other

## Product Information:

Product: 7410-7110-6510-6810-7510  
 9510- Drug Tester 5000  
 Description: A - B - Plus - Screener - Demo  
 Serial #: AR WE - 0080  
 Printer Ser #: AR  
 Whole Instrument \_\_\_\_\_ Top 1/2 \_\_\_\_\_ Sim Ser #: \_\_\_\_\_  
 Other \_\_\_\_\_ Probe Ser # DD P  
**ACCESSORIES:**  
 110 V A/C Adapter  Regulator  Mag Card Rdr# \_\_\_\_\_  
 Printer Paper  Printer Ribbon  Casio # \_\_\_\_\_  
 Mouthpieces  Carrying Case  Dry Gas  
 Other (Please Specify) \_\_\_\_\_ Warranty Exp. Date \_\_\_\_\_

## Repair Information:

Test #

Reason for Return:			
<u>FC</u>			
Part Number	Description	Qty	Total Cost
<u>6808455</u>	<u>FC</u>	<u>1</u>	<u>w/c w</u>
<u>mp cal 71</u>	<u>CAL</u>	<u>1</u>	<u>w/c w</u>
<u>MP Labor</u>	<u>Labor</u>	<u>.5</u>	<u>w/c w</u>

Repair Notes:

FC # ARBL-1887

CAL w/Qc & Ops Check

Service Technician: \_\_\_\_\_

Date: 1/19/2011