

# Dräger safety

## Return & Repair Form

Customer Information:

CUSTOMER New Jersey State Police Cust # \_\_\_\_\_  
 Date of Return: 12/13/06 Carrier: UPS Method: 2nd Day

Product Information:

Product : 7110 7410 6510 6810 Serial # ARWE-0082  
 Description: A B PLUS SCREENER \_\_\_\_\_ Printer Serial # \_\_\_\_\_  
 Whole Inst. Top Half Simulator Serial # \_\_\_\_\_  
 OTHER : \_\_\_\_\_ Probe Serial # \_\_\_\_\_  
 Demo Unit  Returned to stock on: \_\_\_\_\_  
 Accessories: (check all that apply):  
 110VAC Adapter  Regulator  Mag Card Rdr # \_\_\_\_\_  
 Printer Paper  Printer Ribbon  Casio # \_\_\_\_\_  
 Mouthpieces  Carrying Case  Dry Gas  
 Other (specify): \_\_\_\_\_  
 Warranty Expiration Date: \_\_\_\_\_

Repair Information:

Test # \_\_\_\_\_

Reason for Return: Repair See attached (Low EC Readings)

Part #	Description	Qty	Total
<u>W08455</u>	<u>Fuel Cell</u>	<u>1</u>	
<u>MPLA302</u>	<u>LABOR</u>	<u>.5</u>	
<u>MPCAL71</u>	<u>7110 CAL</u>	<u>1</u>	

Repair Notes: Replaced questionable P/C. Found the EC sample tube crimped and replaced it. Full Cal ops check + Q.C.

CS

12/14/06