

# Return & Repair Form:

## Customer Information:

Company Name: UNION COUNTY PD B: \_\_\_\_\_ S: \_\_\_\_\_  
 Date Received: 05-18-2009 Date Given to Service: 05-18-2009  
 Carrier: FedEx - UPS - DHL - USPS Method: GRD - NDA - 2DY - 3DY Other

## Product Information:

Product: 7410 7110 - 6510 - 6810 Serial #: AR WE - 0083  
 Description: A - B - Plus - Screener - Demo Printer Ser #: AR - \_\_\_\_\_  
 Whole Instrument Top 1/2 Sim Ser #: \_\_\_\_\_  
 Other \_\_\_\_\_ Probe Ser # DD P - \_\_\_\_\_  
**ACCESSORIES:**  
 110 V A/C Adapter  Regulator  Mag Card Rdr# \_\_\_\_\_  
 Printer Paper  Printer Ribbon  Casio # \_\_\_\_\_  
 Mouthpieces  Carrying Case  Dry Gas  
 Other (Please Specify) \_\_\_\_\_ Warranty Exp. Date \_\_\_\_\_

## Repair Information:

FC

Test #

Reason for Return:

Part Number	Description	Qty	Total Cost
6808455	FC	1	N/C WARR
6808486	PLATE	1	N/C WARR
MP CAL 71	CAL	1	N/C WARR
MP Labor	Labor	.5	N/C WARR

Repair Notes:

REPLACED FC + PLATE  
CAL w/DC & OPS check

Service Technician: [Signature]

Date: 05-19-2009