

Return & Repair Form: _____

Customer Information:

Company Name: Hammerton B: _____ S: _____
 Date Received: 1-13-09 Date Given to Service: _____
 Carrier: (FedEx) - UPS - DHL - USPS Method: GRD - NDA - 2DY - (3DY) - Other

Product Information:

Product: 7410 (7110) - 6510 - 6810 Serial #: AR W F - 0349
 Description: A - B - Plus - Screener - Demo Printer Ser #: AR -
 Whole Instrument Top ½ Sim Ser #: _____
 Other _____ Probe Ser # DD P -

ACCESSORIES:

110 V A/C Adapter Regulator Mag Card Rdr# _____
 Printer Paper Printer Ribbon Casio # _____
 Mouthpieces Carrying Case Dry Gas
 Other (Please Specify) _____ Warranty Exp. Date _____

Repair Information:

Test #

Reason for Return: FC / HIGH IR / PURGE ERRORS.


Part Number	Description	Qty	Total Cost
<u>6808455</u>	<u>FC</u>	<u>1</u>	<u>N/C WORK</u>
<u>13336</u>	<u>TUBC</u>	<u>9"</u>	<u>N/C WORK</u>
<u>MPCAC 71</u>	<u>CAC</u>	<u>1</u>	<u>N/C WORK</u>
<u>MPLMBOR</u>		<u>.5</u>	<u>N/C WORK</u>

Repair Notes:

PINCH IN PUMP TUBE - REPLACED - LOOSE PUMP WIRE - CORRECTED
DEBRIS IN COUETTE - BLEW OUT W/ AIR

CAC W/ QC & OPS CHECKS

Service Technician: _____



Date: 01/15/2009